



A STUDY ON MAGNETIC RESONANCE (MR) SAFETY: ESSENTIAL CONSIDERATIONS FOR PATIENT SAFETY

Vanshika

Tutor, Adesh Institute of Allied & Healthcare Professions, Adesh University, Bathinda, Punjab-151001, India

ABSTRACT

Magnetic Resonance Imaging (MRI) is an advanced diagnostic tool esteemed for its high-resolution imaging capabilities without the application of ionizing radiation. However, despite its non-invasive nature, the distinctive physical properties of MRI, which include strong static magnetic fields, rapidly changing gradient magnetic fields, and radiofrequency (RF) fields, introduce unique safety challenges. This study aims to present a comprehensive overview of critical considerations for ensuring patient safety in the MR environment. Key areas of focus include the risks posed by ferromagnetic projectiles due to the static magnetic field, potential thermal effects and burns from RF energy deposition (Specific Absorption Rate - SAR), hearing damage from acoustic noise, and the safe management of patients with implanted medical devices (e.g., pacemakers, neurostimulator, aneurysm clips) or foreign bodies. The abstract will emphasize the crucial role of stringent safety protocols, including thorough patient screening, the establishment of clearly defined MR safety zones, continuous patient monitoring, and adherence to established guidelines from organizations such as the American College of Radiology (ACR) and the International Commission on Non-Ionizing Radiation Protection (ICNIRP). Additionally, the importance of a well-trained MR safety team, consisting of the MR Medical Director (MRMD), MR Safety Officer (MRSO), and MR Safety Expert (MRSE), in fostering a culture of safety and minimizing adverse events will be highlighted. By addressing these vital considerations, healthcare facilities can significantly enhance patient well-being and optimize the safe use of MRI technology.

KEYWORDS : MR Safety, Patient Safety, Magnetic Resonance Imaging (MRI), Magnetic Fields (Static magnetic field, Gradient magnetic fields, Radiofrequency (RF) fields)

I. INTRODUCTION

Magnetic Resonance (MR) imaging is a significant diagnostic tool, but the unique environment it entails brings forth specific safety challenges that call for strict protocols to ensure the safety of patients. A comprehensive grasp of these critical elements is important for all healthcare workers involved in MR procedures.

Here's a review of the essential aspects of MR safety for patient safety:

1. Understanding the MR Environment and its Hazards:

- **Powerful Static Magnetic Field (B0):** The primary and most significant hazard. The magnet is always on, even when not scanning. It can turn ferromagnetic objects (even small ones like hairpins, coins, or oxygen tanks) into dangerous projectiles, posing a severe risk of injury or death to patients and staff.
- **Gradient Magnetic Fields (dB/dt):** These rapidly switching magnetic fields are responsible for spatial encoding of the MRI signal. They generate:
 - **Loud Acoustic Noise:** Can reach 100 dB or more, requiring mandatory hearing protection (earplugs or headphones) for patients and anyone in the scanner room.
 - **Peripheral Nerve Stimulation (PNS):** Patients may experience tingling, twitching, or muscle contractions, usually transient and harmless.
- **Radiofrequency (RF) Fields (B1):** Used to generate the MR signal, these fields can cause:
 - **Tissue Heating:** Can lead to burns, especially in patients with impaired thermoregulation, those in casts, or with certain implants. Proper patient preparation (removing metallic objects, using insulation) is vital.
 - **Cryogen Hazards (e.g., Liquid Helium):** Superconducting magnets are cooled by cryogenics. An accidental "quench" (rapid loss of superconductivity) can rapidly release large volumes of cold gas, displacing oxygen and posing an asphyxiation risk. MRI rooms should have proper ventilation and oxygen monitors.
 - **Claustrophobia and Anxiety:** The confined space and loud noises can induce significant anxiety in patients. Open MRI machines, sedation options, and clear communication can help mitigate this.

2. MR Safety Zones:

Most MR facilities follow a four-zone system, as outlined by organizations like the American College of Radiology (ACR), to control access and manage risks:

- **Zone I:** Public access areas (e.g., waiting rooms, reception) where the magnetic field poses no hazard.
- **Zone II:** Areas where patients are screened and prepared (e.g., waiting rooms, dressing rooms). Patients are under general supervision.
- **Zone III:** Restricted, controlled-access areas immediately

adjacent to the MRI scanner room (e.g., control room). Only approved MR personnel and screened patients are allowed.

- **Zone IV:** The MRI scanner room itself, posing the highest risk. Access is strictly controlled, usually through Zone III.

3. Comprehensive Patient Screening:

Thorough screening is the cornerstone of MR patient safety. This involves:

- **Detailed Medical History:** Identifying any implanted devices, previous surgeries, or medical conditions.
- **Implant and Device Compatibility:** Crucial for preventing serious adverse events. Devices are categorized as:
 - **MR Safe:** Non-metallic, non-conductive, non-magnetic. Safe in all MR environments.
 - **MR Conditional:** Safe under defined conditions (e.g., specific field strength, SAR limits). Manufacturer's guidelines must be strictly followed.
 - **MR Unsafe:** Poses an unacceptable risk in the MR environment.
- Patients with pacemakers, defibrillators, certain aneurysm clips, cochlear implants, or metallic foreign bodies (especially in the eyes) are often absolute contraindications unless the device is specifically labeled MR Conditional and the scan parameters are carefully managed.
- **Pregnancy Status:** While MRI is generally considered safe during pregnancy, contrast agents are typically avoided unless absolutely necessary, and first-trimester scans are usually postponed unless urgent.
- **Allergies to Contrast Agents:** Especially for gadolinium-based contrast agents, a thorough allergy history is essential.
- **Ability to Cooperate:** Patients must be able to remain still for the duration of the scan. Sedation may be considered for anxious or pediatric patients.
- **Removal of All Metallic Objects:** Jewelry, watches, cell phones, hairpins, and clothing with metallic threads or fasteners must be removed.

4. Personnel and Training:

- **Trained MR Personnel:** All individuals working in the MR environment must undergo comprehensive training, typically categorized into Level I and Level II personnel, with increasing levels of responsibility and knowledge.
- **Designated Roles:** Facilities should establish clear roles such as MR Medical Director (MRMD), MR Safety Officer (MRSO), and MR Safety Expert (MRSE) to oversee and enforce safety policies.
- **Ongoing Education:** The rapidly evolving field of MRI necessitates continuous education and updates on safety guidelines and best practices.

5. Emergency Preparedness:

- **Emergency Procedures:** Clear protocols for emergencies like

cardiac arrest, fire, or magnet quench are essential. This includes knowing how to safely remove a patient from the magnet and where to access MR-compatible emergency equipment.

- **Quench Pipe Inspection:** Regular inspection of quench pipes is vital to ensure safe venting of cryogens during an emergency.

6. Best Practices and Continuous Improvement:

- **Adherence to Guidelines:** Facilities should strictly follow established guidelines from organizations like the ACR and national regulatory bodies.
- **Clear Communication:** Effective communication between patients, referring physicians, and MR staff is paramount to ensure all safety concerns are addressed.
- **Patient Education:** Educating patients about the MRI procedure, potential risks, and safety measures significantly contributes to their cooperation and reduces the likelihood of complications.

7. Facility Design and Signage:

- Properly marked MRI zones and clear warning signs indicating the presence of the strong magnetic field are fundamental.
- Restricted access (e.g., keycard entry) to Zone III and IV is crucial.

8. Specific Considerations:

- **Claustrophobia/Anxiety:** Patients should be informed about the scan environment. Sedation or "open MRI" options may be available.
- **Children:** Special considerations for sedation, monitoring, and communication.
- **Contrast Agents:** Guidelines for gadolinium-based contrast agents, including kidney function assessment and risks of retention.

II. Patient Monitoring and Emergency in a Magnetic Resonance Imaging:

Monitoring devices like pulse oximetry, ventilators are now available as MR compatible devices that can be used safely in scanner room. These should be kept as away from magnet bore as possible. In spite of availability of these devices the first approach in case of emergency must be shifting the patient out of scanning room as early as possible and start resuscitation.

III. Precautions:

1. Always screen the patient and accompanying person for any metallic objects. Metallic objects can form projectile because of strong magnetic attraction. This can lead to life threatening consequences.
2. Always see to it that wires and coils are well insulated and are not touching patient's body. It can cause burns. Patient's body part should also not be touching magnet bore.
3. Avoid loop formation: - It can lead to induction of current and burns.
4. In case of emergency first approach must be to remove the patient out of scanner room as early as possible and start resuscitation.
5. Doors of scanner room should have label with pictures of object that are strictly prohibited to take inside scanner room.

IV. Safety Instruction and Emergency Protocols in a Magnetic Resonance Environment:

MR personnel must be informed about the possible adverse effects of MRI. It is necessary to complete MRI safety training annually, which should cover both the technical and medical foundations of MRI. To comprehend the risks associated with MR, training programs ought to include practical demonstrations of the effects of ferromagnetic materials. Another significant element of the training program is the screening of patients to identify any ferromagnetic objects, implants, allergies to contrast agents, kidney conditions, and other relevant medical history. Understanding this information is crucial for preventing injuries. Since most MR complications are related to burns, an MRI safety course should address methods to minimize the risk of thermal burns. The most effective way to mitigate the risk of burns is to require patients to wear MR-safe clothing. Some clothing may have small amounts of metallic fibers, which are often noted on the label of the garment. Additionally, MR personnel should be aware of the hearing risks linked to MR. During the scan, the gradient system can generate very high levels of acoustic noise. Patients need to wear earplugs and headphones to prevent hearing damage. As severe allergic reactions to gadolinium chelates can occur, MR personnel must be trained to respond to these emergencies.

V. CONCLUSION:

Magnetic Resonance (MR) is a secure, non-invasive diagnostic method that is highly beneficial for numerous clinical applications. It is vital to understand the safety protocols and procedures to ensure the safe implementation of MR for patients and research subjects. Segmenting the MR area into four zones, each with different restrictions, is essential for effective organization. In addition, adequate screening is necessary to assess a patient's risk within the MR unit. Avoiding ferromagnetic materials is crucial to prevent injuries and protect MR equipment. MR staff must be aware of the safety risks associated with MRI, which include projectile injuries, high Specific Absorption Rate (SAR), thermal burns, and the side effects of Gadolinium-Based Contrast Agents (GBCAs). Proper organization and screening in the MR unit, along with thorough training for MR personnel, are important steps to ensure the safe application of MR.

Keywords

MR Safety, Patient Safety, Magnetic Resonance Imaging (MRI), Magnetic Fields (Static magnetic field, Gradient magnetic fields, Radiofrequency (RF) fields), Hazards (Projectile effect, Burns, Acoustic noise, Heating, Implant/device malfunction), Risk Management, Safety Protocols, Guidelines (e.g., ACR Guidance Document on MR Safety), Screening (Patient screening, Personnel screening), MR Zones (Zone I, Zone II, Zone III, Zone IV), Implants, Medical Devices (MR Safe, MR Conditional, MR Unsafe), Ferromagnetic Objects, Cryogenics (e.g., Liquid Helium, Quench), Informed Consent, Radiology, Medical Physics, Quality Improvement, Diagnostic Imaging.

VI. REFERENCES

1. Recommended responsibilities for management of MR safety. Calamante F, Ittermann B, Kanal E, Norris D. <https://onlinelibrary.wiley.com/doi/pdf/10.1002/jmri.25282>. *J MagnReson Imaging*. 2016; 44:1067–1069. doi: 10.1002/jmri.25282. [DOI] [PubMed] [Google Scholar]
2. Magnetic resonance safety. Sammet S. *AbdomRadiol (NY)* 2016;41:444–451. doi: 10.1007/s00261-016-0680-4. [DOI] [PMC free article] [PubMed] [Google Scholar]
3. ACR manual on MR safety 2020. <https://www.acr.org/-/media/ACR/Files/Radiology-Safety/MR-Safety/Manual-on-MR-Safety.pdf>
4. Implementation of a comprehensive MR safety course for medical students. Sammet S, Sammet CL. *J MagnReson Imaging*. 2015; 42:1478–1486. doi: 10.1002/jmri.24993. [DOI] [PMC free article] [PubMed] [Google Scholar]
5. The dangers of fabric in MRI. Styan T, Hoff M. *CurrProblDiagnRadiol*. 2023; 52:6–9. doi: 10.1067/j.cpradiol.2022.07.011. [DOI] [PubMed] [Google Scholar]
6. American College of Radiology. ACR manual on MR safety, 2020. Available from https://www.acr.org/-/media/ACR/Files/Radiology_Safety/MR_Safety/Manual.on.MR_Safety.pdf (Accessed April 30, 2021).
7. Hoff M, McKinney A, Shellock F, et al. Safety considerations of 7.T MRI in clinical practice. *Radiology* 2019; 292: 509–18. [DOI] [PubMed]
8. http://www.simplyphysics.com/mri_safety.html.
9. Kim SJ, Kim KA. Safety issues and updates under MR environments. *Eur J Radiol* 2017; 89: 7–13. [DOI] [PubMed] [Google Scholar]
10. Keevil S. Safety in magnetic resonance imaging. *Int J Med Phys* 2016; 4: 26–34. [Google Scholar]