



A COMPARATIVE ANALYSIS OF VARIOUS INTRAOCULAR LENS POWER CALCULATION FORMULAE TO ACHIEVE EMMETROPIA FOLLOWING CATARACT SURGERY BY PHACOEMULSIFICATION

OPHTHALMOLOGY

Dr. Rakesh Maggon

Hod, Dept. Of Ophthalmology, Command Hospital, Kolkata, West Bengal.

Dr. Sumit Kumar Singh*

Dept. Of Ophthalmology, Command Hospital, Kolkata, West Bengal. *Corresponding Author

Dr. Maneesh Jha

Dept. Of Ophthalmology, Command Hospital, Kolkata, West Bengal.

Dr. Sandeep Gupta

Dept. Of Ophthalmology, Command Hospital, Kolkata, West Bengal.

ABSTRACT

For the purpose of study the eyes are divided into following subgroups depending on the formulae used

- SRK/T
- HOFFER Q
- HOLLADAY
- HAIGIS

This is a comparative analysis of various intraocular lens power calculation formulae to achieve emmetropia following Cataract surgery by phacoemulsification.

We found in our study a total of 240 eyes were studied and subdivided into 4 groups as per IOL Formulae at Department of Ophthalmology, Command Hospital (Eastern Command), Kolkata between January 2018 – June 2019 were taken for the study. 60 Patients were randomly assigned to 4 groups each were underwent PKE with IOL implantation by a single surgeon using one of the 4 formulae.

In our study distribution of mean AL in the groups was statistically significant ($p=0.0099$). Distribution of mean residual SRE with in the groups was statistically significant ($p<0.0001$).

It was found that difference of mean small eye and normal eye with in groups was statistically significant but difference of mean large eye with in groups was not statistically significant.

We found that aver all in all eyes Hoffers Q Formula is coming as most accurate. In Small eyes as well as in Normal eyes Hoffer Q IOL formula is most accurate followed by Holladay formula, SRK-T formula and Haigis formula and In case of Large eyes Holladay formula is coming better followed by Hoffer Q, Haigis and SRK-T formula.

KEYWORDS

Intraocular Lens Power, Formulae, Cataract Surgery, Phacoemulsification

INTRODUCTION

Cataract affects nearly 9–12 million Indians annually after the age of 50 years' which exceeds glaucoma, macular degeneration, and diabetic retinopathy numbers combined. Cataract surgery is the most frequently performed surgery in India, with a success rate of nearly 98% and excellent visual rehabilitation.^{1,2}

Formulae for calculation of IOL power have evolved since Harold Ridley implanted the first IOL in 1949 and was surprised to get a postoperative refraction of $-18\text{DS}/+6\text{DC}/20^\circ$.³ He implanted an IOL of power 23D since that was the power of the natural lens. Since then the ophthalmologists have strived to arrive at a reliable IOL formula. The evolution of IOL formulae can be classified on the basis of their derivation as (a) theoretical formulae, (b) those based on regression analysis, and (c) a combination of both principles. Various phases of IOL power Formulae evolution have been labelled as various generations. Initially, a power estimation method was used, and the patient's preoperative refraction was taken into account. Later, this evolved into a more specific calculation based on biometric parameters. First-generation formulae, such as SRK (Sanders, Retzlaff and Kraf) were based on the regression to estimate power of the lens which was based on cornea (K), axial length (AL), and A-Constant for the specific IOL to estimate power for IOL. They were found to be accurate in average sized eyes, but for shorter or longer eyes these formulae were not reliable. In second generation formulae, SRK II, a modification was made.⁴

In order to achieve emmetropia after cataract surgery, choosing from this plethora of IOL power formulae remains a challenge. For selection of IOL before cataract surgery the anatomical and optical parameters of eye have to be considered. In most of the cases we aim to achieve emmetropia but in some cases, depending on the specific needs of the individual patients, ametropia may result and some residual myopia may be required.

In calculating IOL power, one of the most important parameters is ocular AL. AL is a combination of anterior chamber depth (ACD), lens

thickness and vitreous cavity depth. It can change the IOL power by up to 2.5 to 3 Dioptres (D). Another important parameter is the corneal curvature and keratometry (K) is the measurement of the corneal radius of curvature. The cornea is the transparent part of the eye that covers the iris, pupil and anterior chamber and it accounts for around two thirds of the eye's total refractive power. Changes in corneal refractive power can alter the IOL power in a ratio of nearly 1:1. Other than AL and K, other parameters that may be also required depending on the type of formula used are the preoperative ACD and the corneal white-to-white distance (WTW; also called the horizontal corneal diameter). The anterior chamber is the fluid-filled space between the iris and the innermost surface of the cornea and the WTW distance is the horizontal distance between the borders of the corneal limbus.⁵

The manufacturer IOL constant (A-constant) as available in the software of the IOL master for different types of IOLs was used. IOL power calculation will be performed in all eyes using any of the four formulae SRK-T, HOFFER Q, HAIGIS' and HOLLADAY'S.

Postoperatively, all patients was examined after six week for the purpose of study and a manual refraction was performed by a single investigator at 1 mtr distance using streak retinoscope and confirmed with Autorefracto-keratometer for all patients. The refractionist was blinded to the group to which patients belonged. Residual refractive power so obtained was then converted into spherical equivalent (SE) for the analysis.

For the purpose of study the eyes are divided into following subgroups depending on the formulae used

- SRK/T
- HOFFER Q
- HOLLADAY
- HAIGIS

This is a comparative analysis of various intraocular lens power calculation formulae to achieve emmetropia following Cataract surgery by phacoemulsification.

MATERIALS AND METHODS

Patients were undergoing cataract surgery at this centre Department of Ophthalmology, Command Hospital (Eastern Command), Kolkata from January 2018–June 2019.

Study Technique

After obtaining clearance from ethical committee of the institute, the study was conducted in the Department of Ophthalmology, Command Hospital (EC) Kolkata. All patients undergoing cataract surgery by PKE in this centre were taken and an IOL power calculation was done using Zeiss IOL master utilising parameters; AXL, anterior chamber depth, keratometry and lens thickness. 60 Patients were randomly assigned to 4 groups each were underwent PKE with IOL implantation by a single surgeon using one of the 4 formulae.

Postoperative refraction for all 4 groups of patients implanted with IOL using four different formulae at 6 weeks was taken for the purpose of this study. An optimized lens constant and an IOL with a standardised —Allconstant was used for all formulae. Results were also analysed according to AL, with eyes classified as Short Eyes (<22mm), Normal Eyes (22–24 mm), and Long Eyes (>24mm)

PLAN FOR ANALYSIS OF DATA

For statistical analysis data were entered into a Microsoft excel spreadsheet and then analyzed by SPSS (version 24.0; SPSS Inc., Chicago, IL, USA) and GraphPad Prism version 5. Data had been summarized as mean and standard deviation for numerical variables and count and percentages for categorical variables. Two-sample t-tests for a difference in mean involved independent samples or unpaired samples. Without other qualification, 'chi-squared test' often is used as short for Pearson's chi-squared test. Unpaired proportions were compared by Chi-square test or Fischer's exact test, as appropriate. p -value ≤ 0.05 was considered for statistically significant.

RESULTS AND ANALYSIS

We found in our study a total of 240 eyes were studied and subdivided into 4 groups as per IOL Formulae at Department of Ophthalmology, Command Hospital (Eastern Command), Kolkata between January 2018–June 2019 were taken for the study. In this study Haigis groups 13 (21.7%) patients were female and 47 (78.3%) patients were male, Hoffer Q groups 19 (31.7%) patients were female and 41 (68.3%) patients were male. In Holladay group, 24 (40.0%) patients were female and 36 (60.0%) patients were male. In SRK-T, 3 (5.0%) patients were female and 57 (95.0%) patients were male. Association of gender within the subgroups was statistically significant ($p=0.0001$).

Our study showed that in HAIGIS group the mean age of the patients was 61.98 ± 8.51 years. In HOFFERQ group the mean age of the patients was 62.85 ± 7.88 years. In HOLLADAY group the mean age of the patients was 62.56 ± 7.24 years. In SRK-T group the mean age of the patients was 61.76 ± 9.43 years. Distribution of mean age with in groups was not statistically significant ($p=0.8827$).

In our study we showed that HAIGIS group the mean Axial length of the patients was 22.96 ± 0.62 . In HOFFERQ group the mean AL of the patients was 22.93 ± 0.62 . In HOLLADAY group, the mean AL of the patients was 23.20 ± 0.62 . In SRK-T group the mean AL of the patients was 23.27 ± 0.76 . Distribution of mean AL in the groups was statistically significant ($p=0.0099$).

We found in HAIGIS group the mean AL group of the patients was 1.98 ± 0.2249 . In HOFFER Q group the mean AL group of the patients was 2.00 ± 0.26 . In HOLLADAY group the mean AL group of the patients was 2.05 ± 0.38 . In SRK-T group the mean AL group of the patients was 2.15 ± 0.44 . Distribution of mean AL group with in. groups was statistically significant ($p=0.0358$).

In our study we showed HAIGIS group the mean refractive error of the patients was -0.33 ± 0.80 . In HOFFER Q group the mean refractive error of the patients was 0.05 ± 0.78 . In HOLLADAY group the mean refractive error of the patients was 0.21 ± 0.73 . In SRK-T group the mean refractive error of the patients was 0.39 ± 0.62 . Distribution of mean residual SRE with in the groups was statistically significant ($p<0.0001$).

Our study showed that in HAIGIS group the mean residual SRE small eye of the patients was 0.68 ± 0.26 . In HOFFERQ group, the mean residual SRE small eye of the patients was -0.62 ± 0.17 . In HOLLADAY group the mean residual SRE small eye of the patients

was 0.83 ± 0.14 . In SRK-T group the mean residual SRE small eye of the patients was 0.75 ± 0.00 . Distribution of mean small eye with in groups was statistically significant ($p=0.0008$). [Table 6 & Fig. 06] In HAIGIS group the mean residual SRE in normal eye of the patients was -0.36 ± 0.80 . In HOFFERQ group, the mean residual SRE in normal eye of the patients was 0.07 ± 0.80 . In HOLLADAY group the mean residual SRE in normal eye of the patients was 0.19 ± 0.75 . In SRK-T group the mean residual SRE in normal eye of the patients was 0.31 ± 0.63 . Distribution of mean normal eye with in groups was statistically significant ($p<0.0001$).

In our study we showed that HAIGIS group the mean residual SRE in large eye of the patients was -0.50 ± 0.00 . In HOFFER group the mean spherical residual refractive error in large eye of the patients was 0.25 ± 0.00 . In HOLLADAY group the mean spherical residual refractive error in large eye of the patients was 0.12 ± 0.64 . In SRK-T group the mean residual spherical refractive error in large eye of the patients was 0.68 ± 0.54 . Distribution of mean large eye with in groups was not statistically significant ($p=0.1189$).

DISCUSSION

Cataract surgery is a kerato-refractive procedure where the patient expects excellent visual rehabilitation in the form of improved vision, contrast sensitivity and depth perception. The most important cause of a refractive surprise after IOL implantation is an incorrect IOL power calculation which is the single most important determinant of patient satisfaction after surgery. It depends on multiple variables such as AL, K, ACD, and ELP.

A prospective study was done by to compare the accuracy of 4 formulae by calculating the difference between predicted and the final refraction after PKE using each formula for IOL power calculation. PCI-based Zeiss IOL master was used to determine AL, anterior chamber depth, keratometry, lens thickness for pre-operative IOL power calculation as also postoperative refraction for all the groups of patients implanted with IOL using four different formulae.

Biometric data was entered into each of IOL power calculation formulae and results was compared to final refraction after 6 weeks. The following 04 formulae were used in this study for calculation of IOL Power i.e SRK-T, Holladay, Hoffer-Q and Haigis. Total 240 eyes were divided into four groups of 60 eyes each.

In this study, in Haigis group 13 (21.7%) patients were female and 47 (78.3%) patients were male, in Hoffer Q group 19 (31.7%) patients were female and 41 (68.3%) patients were male. In Holladay group, 24 (40.0%) patients were female and 36 (60.0%) patients were male. In SRK-T, 3 (5.0%) patients were female and 57 (95.0%) patients were male. It has been found that gender distribution in all the subgroups showed male predominance which was statistically significant all groups ($p=0.0001$).

In Haigis group, mean AL of the patients was 22.96 ± 0.62 , in Hoffer Q group mean AL patients was 22.93 ± 0.62 , in Holladay group mean AL of the patients was 23.20 ± 0.62 and in SRK-T group mean AL of the patients was 23.27 ± 0.76 . Mean AL was marginally higher in Holladay and SRK-T group. This was statistically significant ($p=0.0099$). [Table 3 & Fig.3] Also in Present study we found that in Haigis group, the mean ACD of the patients was 1.98 ± 0.22 . In Hoffer Q group, the mean ACD of the patients was 2.00 ± 0.26 . In Holladay group, the mean ACD of the patients was 2.05 ± 0.38 . In SRK-T group, the mean ACD of the patients was 2.15 ± 0.44 . Mean ACD was higher in SRK-T group in comparison to other groups and this was statistically significant ($p=0.0358$).

In our study Gender, AL and ACD distributions with in the subgroups showed statistically significant difference in spite of adequate randomization. This may have bearing on the final results.

In Haigis group, the mean residual Spherical Refractive Error (SRE) of the patients was -0.33 ± 0.80 , in Hoffer Q group residual SRE was 0.05 ± 0.78 , in Holladay group mean residual SRE was 0.21 ± 0.73 and in SRK-T group mean residual SRE was 0.39 ± 0.62 . The lowest mean residual SRE was found in Hoffer Q followed by Holladay, SRK-T and Haigis. This association was statistically significant ($p<0.0001$).

In a study done by **Fam HB et al⁶ (2008)** in 37 eyes of 37 patients IOL power was calculated using the formulae : (SRK), SRK-T, Holladay ,

Hoffer Q, The calculated IOL power was compared and used to determine the mean error and mean absolute error of refractive outcome for each eye. The mean absolute error achieved using the Hoffer Q method was (0.75 $\pm$ 0.52 D), Holladay was (0.75 $\pm$ 0.62 D), in SRK-T it was (0.76 $\pm$ 0.60 D) and with SRK-T formula, 51.4% of eyes were within $\pm$ 0.50 D of emmetropia and 67.6% of eyes were within $\pm$ 1.00 D. In the Holladay group the highest percentage (81.1%) of eyes within $\pm$ 1.00 D and 45.9% of the eyes in this group were within $\pm$ 0.50 D.

In our study least mean residual SRE was found in Hoffer Q (0.05) followed by Holladay (0.21), SRK-T(0.39) and Haigis formula (-0.33). In small eyes (<20mm), in Haigis group, the mean residual SRE was 0.68 $\pm$ 0.26, in Hoffer Q group mean residual SRE was -0.62 $\pm$ 0.17, in Holladay group it was 0.83 $\pm$ 0.14 and in SRK-T group mean residual SRE was 0.75 $\pm$ 0.00. Mean residual SRE in small eyes was higher in Hoffer Q group followed by Holladay, SRK-T and Haigis groups. This difference was statistically significant (p=0.0008).

In a study done by **Gavin EA et al**⁷ IOL power was calculated using SRK-T and Hoffer Q formulae. Refractive outcome was measured and the accuracy of the two formulae were compared in 41 eyes with an AL <22 mm. It was found that the Hoffer Q formula showed a mean prediction error of 0.61 D while SRK-T showed a mean prediction error of 0.87 D. A paired t-test found that the Hoffer Q was significantly more accurate than the SRK-T formula (P<0.001). Our study has also shown a similar result and has found Hoffer Q formula to be more accurate than the SRK-T formula in small eyes. Hoffer Q was more accurate than Holladay, SRK-T and Haigis formula.

Another study done by **Karabela Y et al**⁸ compared the predictability of different IOL power calculation formulae using (SRK) II, SRK-T, Haigis, Hoffer Q in eyes with AL <22 mm using IOL master. Here Haigis formula provided most accurate results & Hoffer Q was also found to be comparable and could be used as an alternative. However, in a study by **Moschos MM et al**⁹ Haigis formula had statistically significant smaller mean residual refraction in comparison to Holladay, Hoffer Q, and SRK/T. The Haigis formula predicted more eyes with residual within $\pm$ 0.5 D and $\pm$ 1.0 D of predicted SRE compared to other formulae. This was in contrast to our study in which Hoffer Q were most accurate with lowest mean refractive error followed by Holladay, SRK-T and Haigis in smaller eyes.

In our study, in normal eyes (AL 20 to 24mm) in Haigis group the mean residual SRE was -0.36 $\pm$ 0.80, in Hoffer Q group mean residual SRE was 0.07 $\pm$ 0.80, in Holladay group it was 0.19 $\pm$ 0.75 and in SRK-T group mean residual SRE was 0.31 $\pm$.63. This was statistically significant (p<0.0001). In normal sized eyes Hoffer Q emerged the most accurate formula followed by Holladay, SRK-T and Haigis.

A study by **Bai L et al**¹⁰ found that Hoffer Q formula appeared to be more accurate when measuring ALs with A-Scan, whereas Haigis formula was more accurate when combined with IOL Master (0.37 $\pm$ 0.14). For selection of IOL formula in eyes with high hyperopia, Haigis formula would be the most accurate using IOL Master analysis, but the Hoffer Q was better when using A-scan. In our study using IOL Master, Hoffer Q formula was the most accurate in Normal eyes followed by Holladay, SRK-T and Haigis and the difference was statistically significant (p<0.0001). Distribution of mean residual refractive error in large eye with in groups was not statistically significant (p=0.1189). In their study **Mansour MN et al**⁹ found that the accuracy of SRK/T and Haigis formulae used for intraocular lens (IOL) power calculation was significant in eyes with an AL of more than 25 mm. The proportion of patients having a prediction error within $\pm$ 0.50 of SRK/T formula (54.29%) was comparable to those of Haigis (55.71%). In another study done by **Doshi D et al**¹¹ it was found that in eyes with AL more than 24.5 mm Hoffer Q, Holladay, SRK/T and Haigis formulae were equally accurate. **Karabela Y et al**¹² found that in eyes $\geq$ 24.6 mm the SRK/T formula performs well and shows good predictability in eyes with long ALs. In our study Holladay formula was found better in long eyes followed by Hoffer Q, Haigis and SRK-T but distribution of mean large eye with in groups was not statistically significant (p=0.1189). In this Prospective study after analysis of results we found that in all eyes Hoffer Q Formula is coming as most accurate.

In Small eyes as well as in Normal eyes Hoffer Q IOL formula is most accurate followed by Holladay formula, SRK-T formula and Haigis formula and In case of Large eyes Holladay formula is coming better followed by Hoffer Q, Haigis and SRK-T formula.

CONCLUSION

At present, cataract surgery is one of the most frequently performed and successful operations in the world. As cataract surgery technology and intraocular lens technology have improved remarkable and become safe; the patients have been expecting better postoperative refractive results, which are determined by the precise intraocular lens power calculation. There are multiple techniques and methods to measure corneal power and AL necessary for different IOL calculation formulae existing at present time. In this Prospective study after analysis of results we found that aver all in all eyes Hoffer Q Formula is coming as most accurate. In Small eyes as well as in Normal eyes Hoffer Q IOL formula is most accurate followed by Holladay formula, SRK-T formula and Haigis formula and In case of Large eyes Holladay formula is coming better followed by Hoffer Q, Haigis and SRK-T formula.

Table 1: Distribution of mean in four groups

		No.	Mean	SD	Min	Max	Median	p-value
Age	HAIGIS	60	61.98	8.51	39.00	80.00	62.00	0.8827
	HOFFER Q	60	62.85	7.88	42.00	76.00	64.00	
	HOLLADAY	60	62.56	7.24	44.00	76.00	62.50	
	SRK-T	60	61.76	9.43	32.00	87.00	62.00	
Axial length	HAIGIS	60	22.96	0.62	21.32	24.05	22.91	0.0099
	HOFFER Q	60	22.93	0.62	20.93	24.11	22.84	
	HOLLADAY	60	23.20	0.62	21.79	24.11	23.31	
	SRK-T	60	23.27	0.76	20.93	24.84	23.25	
Axial length group	HAIGIS	60	1.98	0.22	1.00	3.00	2.00	0.0358
	HOFFER Q	60	2.00	0.26	1.00	3.00	2.00	
	HOLLADAY	60	2.05	0.38	1.00	3.00	2.00	
	SRK-T	60	2.15	0.44	1.00	3.00	2.00	
SRE	HAIGIS	60	0.33	0.80	-1.87	1.75	-0.50	<0.0001
	HOFFER Q	60	0.05	0.78	-1.57	1.50	0.25	
	HOLLADAY	60	0.21	0.73	-1.25	1.75	0.50	
	SRK-T	60	0.39	0.62	-1.25	1.50	0.500	
Small Eyes (<20 mm)	HAIGIS	2	0.68	0.26	0.50	0.87	0.68	0.0008
	HOFFER Q	2	-0.62	0.17	-0.75	-0.50	-0.62	
	HOLLADAY	3	0.83	0.14	0.75	1.00	0.75	
	SRK-T	2	0.75	0.00	0.75	0.75	0.75	
Normal Eyes (20-24mm)	HAIGIS	57	-0.36	0.80	-1.87	1.75	-0.57	<0.0001
	HOFFER Q	56	0.07	0.80	-1.57	1.50	0.25	
	HOLLADAY	51	0.19	0.75	-1.25	1.75	0.50	
	SRK-T	47	0.31	0.63	-1.25	1.50	0.50	
Large Eyes (>24 mm)	HAIGIS	1	-0.50	0.00	-0.50	-0.50	-0.50	0.1189
	HOFFER Q	2	0.25	0.00	0.25	0.25	0.25	
	HOLLADAY	6	0.12	0.64	-0.50	1.00	0.00	
	SRK-T	11	0.68	0.54	-0.50	1.25	1.00	

REFERENCES

- Survey of Blindness- India, NPCB- WHO Report. New Delhi: Ministry of Health & Family Welfare, Government of India; 2015- 2018.
- World Health Organization. The World Health Report, Life in 21st Century – A Vision. Geneva: World Health Organization; 2018.
- Williams HP. Sir Herold Ridley's vision. Br J Ophthalmol 2002;120:1198- 573.
- David J. Sir Herold Ridley & His Fight for Sight, He Changed the World so that we may Better. New Jersey; 2006.
- Gavin EA, Hammond CJ. Intraocular lens power calculation in short eyes. Eye 2008;22:935-8.
- Fam HB, Lim KL. A comparative analysis of intraocular lens power calculation methods after myopic excimer laser surgery. Journal of refractive surgery. 2008 Apr 1;24(4):355-60.
- Gavin EA, Hammond CJ. Intraocular lens power calculation in short eyes. Eye. 2008 Jul;22(7):935.
- Karabela Y, Eliacik M, Kocabora MS, Erdur SK, Baybora H. Predicting the refractive outcome and accuracy of IOL power calculation after phacoemulsification using the SRK/T formula with ultrasound biometry in medium axial lengths. Clinical ophthalmology (Auckland, NZ). 2017;11:1143.
- Moschos MM, Chatziralli IP, Koutsandrea C. Intraocular lens power calculation in eyes with short axial length. Indian journal of ophthalmology. 2014 Jun;62(6):692.
- Bai L, Zhang ZP, Yi GL, Wu WJ, Lin HT, Yan PS. Selection of accurate IOL formula in patients with cataract and high hyperopia. [Zhonghua yan ke za zhi] Chinese journal of ophthalmology. 2008 Dec;44(12):1063-5.
- Doshi D, Limdi P, Parekh N, Gohil N. A comparative study to assess the predictability of different IOL power calculation formulae in eyes of short and long axial length. Journal of clinical and diagnostic research: JCDR. 2017 Jan;11(1):NC01.
- Karabela Y, Eliacik M, Kaya F. Performance of the SRK/T formula using A-Scan ultrasound biometry after phacoemulsification in eyes with short and long axial lengths. BMC ophthalmology. 2016 Dec;16(1):96.